

S2-1.2

Low Power Constant Current Driver For Implantable Electrostimulator Of The Lower Esophageal Sphincter

V. Vidiborschii¹, V. Sontea¹, S. Ungureanu², N. Sipitco² and D. Fosa²

¹ *Technical University of Moldova, Chisinau, Republic of Moldova*

² *State University of Medicine and Pharmacy „Nicolae Testemitanu“, Chisinau, Republic of Moldova*

Tone modulation of lower esophageal sphincters (LES) via electrical stimulation is a novel method of gastroesophageal reflux disease (GERD) treatment. Traditionally for output are commonly used constant current drivers, based on different schematic solutions. This allows delivering same stimulation energy even in case of impedance change during patient movements or electrode contact aging or encapsulation. The aim of our work was to design and implement simple constant current driver with digital current control and ultra-low power consumption. This driver was used in a prototype of implantable LES stimulator (WPLES), which effectiveness was confirmed during animal tests.